

Sample Name : OVA Peptide(257-264)
 Cat. No. : RP10611
 Lot. No. : PE9771606

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	11.615	11471	1885	0.199
2	14.756	2800	720	0.049
3	16.486	23681	3659	0.410
4	17.614	61117	6458	1.059
5	17.793	18326	4884	0.318
6	17.924	5592604	666895	96.897
7	18.675	27308	5154	0.473
8	18.912	23008	3722	0.399
9	19.257	11360	1657	0.197
Total		5771675	695034	100.000