

Sample Name : OVA Peptide(257-264)
 Cat. No. : RP10611
 Lot. No. :PE7721607

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.10	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	5.655	17990	2533	0.764
2	11.675	1446	294	0.061
3	12.026	4116	647	0.175
4	15.487	5273	1069	0.224
5	16.507	8643	1586	0.367
6	16.608	7696	1603	0.327
7	16.765	5640	1269	0.239
8	16.980	2283216	390551	96.932
9	17.478	9504	1830	0.403
10	17.639	2412	597	0.102
11	18.330	1868	246	0.079
12	18.763	7674	1072	0.326
Total		2355477	403300	100.000