

Version 3.0

Update: 02/07/2023

## CERTIFICATE OF ANALYSIS

### Product Information

Product Name	SARS-CoV-2 Surrogate Virus Neutralization Test Kit(Research Use Only)
Cat. No.	L00847-A
Lot No.	A240101
Kit Content	The kit contains 10 components: Capture Plate Positive Control Negative Control HRP conjugated RBD HRP Dilution Buffer Sample Dilution Buffer 20× Wash Solution TMB Solution Stop Solution Plate Sealer
Storage	The product is stable for at least 12 months if stored at 2-8°C

### Quality Control

Test Items	Specifications	Test results	Conclusion
Positive Control	OD450 value < 0.3	0.0832	Passed
Negative Control	OD450 value > 1.0	2.4357	Passed

### Caution

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

Operator by / Date: *Dar sheng fan 01/11/2024*Approved by/ Date: *Juyuan yuam 01/11/2024*

*For laboratory research use only. Direct human use, including taking orally and injection and clinical use are forbidden.*