

Version 1.0  
Update: 02/17/2022

## CERTIFICATE OF ANALYSIS

### Product Information

Product Name	cPass SARS-CoV-2 Neutralization Antibody Detection Kit
Cat. No.	L00847-C
Date of manufacture	11/06/2023
Lot No.	A231101
Kit Content	The kit contains 10 components: Capture Plate Positive Control Negative Control HRP conjugated RBD HRP Dilution Buffer Sample Dilution Buffer 20× Wash Solution TMB Solution Stop Solution Plate Sealer
Storage	The product is stable for at least 12 months if stored at 2-8°C

### Quality Control

Test Items	Specifications	Test results	Conclusion
Positive Control	OD450 value < 0.3	0.0836	Passed
Negative Control	OD450 value > 1.0	2.0475	Passed

### Caution

Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

Operator by / Date: *Shenyuan Dai* 12/06/2023  
Approved by / Date: *Yuanjun Lu* 12/06/2023