

Version 3.0

Update: 10/20/2021

CERTIFICATE OF ANALYSIS

Product Information

| | |
|---------------------|---|
| Product Name | cPass™ SARS-CoV-2 Neutralization Antibody Detection Kit |
| Cat. No. | L00847-C |
| Date of manufacture | 10/12/2021 |
| Lot No. | A210702 |
| Kit Content | The kit contains 10 components: Capture Plate Positive Control Negative Control HRP conjugated RBD HRP Dilution Buffer Sample Dilution Buffer 20× Wash Solution TMB Solution Stop Solution Plate Sealer |
| Storage | The product is stable for at least 12 months if stored at 2-8°C |

Quality Control

| Test Items | Specifications | Test results | Conclusion |
|------------------|-------------------|--------------|------------|
| Positive Control | OD450 value < 0.3 | 0.0701 | Passed |
| Negative Control | OD450 value > 1.0 | 2.0927 | Passed |

Caution

For laboratory or further manufacturing use only. Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

Operator by / Date: *Furong Gao* 12/08/2021
Approved by/ Date: *Kungang Lu* 12/08/2021