

Version 4.0

Update: 4/13/2022

## CERTIFICATE OF ANALYSIS

### Product Information

---

Product Name	cPass SARS-CoV-2 Neutralization Antibody Detection Kit
Cat. No.	L00847
Lot No.	A240301
Kit Content	The kit contains 10 components: Capture Plate Positive Control Negative Control HRP conjugated RBD HRP Dilution Buffer Sample Dilution Buffer 20× Wash Solution TMB Solution Stop Solution Plate Sealer
Storage	The product is stable for at least 9 months if stored at 2-8°C

---

### Quality Control

---

Test Items	Specifications	Test results	Conclusion
Positive Control	OD450 value < 0.3	0.0457	Passed
Negative Control	OD450 value > 1.0	2.4343	Passed

---

### Caution

Not intended for household use. If you have any questions about the Certificate of Analysis, please contact our customer service representative at 1-877-436-7274 (Toll-Free), or 1-732-885-9188.

Operator by / Date: *Sheng fan Dai* 2024.04.09  
Approved by/ Date: *Dandan Yao* 2024.04.09