

Sample Name : CEF20
 Cat No. : RP11781
 Lot No. : PE8631902
 Time Processed : 22:54:35
 Month-Day-Year Processed : 03/01/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump B Conc.	5
25.00	Pumps	Pump B Conc.	65
25.01	Pumps	Pump B Conc.	95
27.00	Pumps	Pump B Conc.	95
27.01	Pumps	Pump B Conc.	5
32.00	Pumps	Pump B Conc.	5
32.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: GK12010012

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	12.600	25227	2582	0.877
2	12.799	18319	1464	0.637
3	13.137	16438	1801	0.572
4	13.429	71660	5460	2.492
5	14.203	2737754	409999	95.213
6	14.557	3286	529	0.114
7	15.087	1005	223	0.035
8	15.193	1707	340	0.059
Total		2875395	422398	100.000