

Sample Name :CEF20
 Cat No :RP11781
 Lot No :E3215821B
 Time Processed :22:59:43
 Month-Day-Year Processed :02/06/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
10.00	Pumps	B.Conc	35
35.00	Pumps	B.Conc	95
37.00	Pumps	B.Conc	95
37.01	Pumps	B.Conc	5
45.00	Pumps	B.Conc	5
45.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: GK11010011

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	13.786	31499	2368	1.160
2	14.050	12440	1341	0.458
3	14.505	2601146	382180	95.830
4	14.905	68052	7887	2.507
5	15.256	1210	217	0.045
Total		2714347	393993	100.000