

Sample Name : LCMV gp33-41  
 Cat No. : RP20257  
 Lot No. : E3755421B  
 Time Processed : 2:18:08 AM  
 Month-Day-Year Processed : 02/03/2021  
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow: 1 ml/min  
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	5
35.00	Pumps	Pump A B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment : GR11010440

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	14.599	7455	873	0.235
2	14.808	58859	8483	1.858
3	15.360	5614	885	0.177
4	15.717	4148	462	0.131
5	16.192	940	222	0.030
6	16.543	3036325	420252	95.849
7	16.873	8906	1050	0.281
8	17.279	3206	494	0.101
9	17.770	2254	293	0.071
10	18.011	1206	257	0.038
11	18.121	4779	613	0.151
12	19.222	5593	582	0.177
13	19.563	28529	2715	0.901
Total		3167816	437181	100.000