

Sample Name : LCMVgp33-41
 Cat No.: RP20257
 Lot No.: PE6841904
 Time Processed : 9:05:39
 Month-Day-Year Processed : 05/03/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.00	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ17010509

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	12.973	27388	4359	1.044
2	14.081	1606	292	0.061
3	14.969	2543716	383550	96.985
4	15.146	37534	9727	1.431
5	15.254	5985	1596	0.228
6	20.714	6569	834	0.250
Total		2622799	400358	100.000