

Sample Name :LL-37, scrambled
 Cat No.: RP20332
 Lot No.:E5295320A
 Time Processed :17:31:07
 Month-Day-Year Processed :01/11/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	15
25.00	Pumps	Pump A B.Conc	75
25.01	Pumps	Pump A B.Conc	95
30.00	Pumps	Pump A B.Conc	95
30.01	Pumps	Pump A B.Conc	15
40.00	Pumps	Pump A B.Conc	15
42.00	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment:ZJ19010324

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	13.219	46594	7345	0.363
2	13.669	27111	4672	0.211
3	15.190	41128	4272	0.321
4	16.301	12512265	1622092	97.521
5	16.826	128610	7170	1.002
6	18.037	74614	11531	0.582
Total		12830323	1657082	100.000