

Sample Name :Cys-TAT(47-57)  
 Cat No. :RP20343  
 Lot No. :E3299421C  
 Time Processed :2:27:20 PM  
 Month-Day-Year Processed :04/06/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>


Time	Module	Command	Value
0.01	Pumps	B.Conc	0
25.00	Pumps	B.Conc	50
25.01	Pumps	B.Conc	95
32.00	Pumps	B.Conc	95
32.01	Pumps	B.Conc	5
40.00	Pumps	B.Conc	5
40.01	Controller	Stop	

<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ19010324

<Chromatogram>



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	7.088	64038	14273	0.447
2	7.443	14113143	1187261	98.603
3	7.672	101260	12887	0.707
4	8.313	34726	5806	0.243
Total		14313166	1220227	100.000