

Sample Name : CEF20  
 Cat No. : RP11781  
 Lot No. : E7141521C  
 Time Processed : 10:54:11 AM  
 Month-Day-Year Processed : 03/30/2021  
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow: 1 ml/min  
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
32.00	Pumps	B.Conc	95
32.01	Pumps	B.Conc	5
40.00	Pumps	B.Conc	5
40.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment : ZJ19010324

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	13.095	45053	4099	2.070
2	13.473	8164	1052	0.375
3	14.224	24612	1733	1.131
4	14.689	2069321	371490	95.066
5	14.995	14190	2459	0.652
6	15.497	8184	1359	0.376
7	16.512	7200	1053	0.331
Total		2176725	383246	100.000