

Sample Name:OVA Peptide (323-339)  
 Cat No.:RP10610  
 Lot No.:E3972320I  
 Time Processed :10:26:24  
 Month-Day-Year Processed :09/24/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
31.00	Pumps	Pump A B.Conc	95
31.01	Pumps	Pump A B.Conc	5
40.00	Pumps	Pump A B.Conc	5
40.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment:ZJ19010324

### <Chromatogram>

mV



### <Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	10.817	2730	1179	0.081
2	10.867	16363	7623	0.485
3	11.021	3267080	466832	96.915
4	11.150	57044	19305	1.692
5	11.283	8865	1622	0.263
6	11.664	11651	1982	0.346
7	12.257	7356	1115	0.218
Total		3371088	499657	100.000