

Sample Name :CEF20
 Cat No. :RP11781
 Lot No. :E9414221D
 Time Processed :10:39:03 PM
 Month-Day-Year Processed :04/26/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	5
35.00	Pumps	Pump A B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	15.473	81141	14476	1.494
2	16.076	15324	2706	0.282
3	16.570	98635	6470	1.816
4	17.065	5165827	836327	95.116
5	17.381	44848	7421	0.826
6	17.998	25300	3621	0.466
Total		5431075	871020	100.000