

Sample Name :Adrenomedullin (AM) (1-52), human  
 Cat. No. : RP11288  
 Lot. No. : E0679121A  
 Time Processed : 20:19:31  
 Month-Day-Year Processed : 03/18/2021


Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
31.00	Pumps	Pump A B.Conc	95
31.01	Pumps	Pump A B.Conc	5
40.00	Pumps	Pump A B.Conc	5
40.01	Controller	Stop	

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment: GK11010017



1 Detector A Channel 1 / 220nm

Peak Table

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	15.846	114826	31890	0.683
2	16.049	16138790	1431312	96.042
3	16.272	500959	73487	2.981
4	16.698	11753	2104	0.070
5	17.324	37518	6433	0.223
Total		16803847	1545225	100.000