

Sample Information

Sample Name : CEF20
 Cat. No : RP11781
 Lot. No. : PE0101806

Pump A : 0.065% trifluoroacetic in 100% waters (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)


<<Pump>>
 Flow : 1.0000 mL/min

<<Detector A>>
 Wavelength Ch1 : 220 nm

<<LC Time Program>>

Time	Module	Command	Value	Comment
0.01	Pumps	Solvent B Conc.	5	
25.00	Pumps	Solvent B Conc.	65	
25.01	Pumps	Solvent B Conc.	95	
27.00	Pumps	Solvent B Conc.	95	
27.01	Pumps	Solvent B Conc.	5	
35.00	Pumps	Solvent B Conc.	5	
35.01	Controller	Stop		

<<Column Performance>>
 <Detector A>
 Column : Inertsil ODS-3 4.6 x 250 mm



Peak Table

Peak#	Ret. Time	Height	Area	Height%	Area%
1	13.702	1552	13542	0.516	0.770
2	13.946	609	2873	0.203	0.163
3	14.395	1395	9643	0.464	0.549
4	14.642	684	4233	0.227	0.241
5	14.946	288031	1676397	95.763	95.356
6	15.237	2444	17922	0.813	1.019
7	15.599	1208	8104	0.402	0.461
8	15.776	4851	25332	1.613	1.441
Total		300774	1758047	100.000	100.000