

Sample Name :OVA Peptide (257-264)

Cat. No. : RP10611 Lot. No. : PE6601706

Pump A : 0.065% trifluoroacetic in 100% water (v/v)

Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)

Total Flow:1 ml/min

Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	5	
25.00	Pumps	Pump B Conc.	65	
25.01	Pumps	Pump B Conc.	95	
27.00	Pumps	Pump B Conc.	95	
27.01	Pumps	Pump B Conc.	5	
35.00	Pumps	Pump B Conc.	5	
35.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : AlltimaTM C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	12.986	2315	441	0.248
2	15.788	922	251	0.099
3	15.943	913619	162471	97.956
4	16.117	5826	1591	0.625
5	16.273	2689	393	0.288
6	16.404	4687	894	0.502
7	16.614	1221	237	0.131
8	16.768	299	93	0.032
9	20.060	1101	224	0.118
Total		932680	166596	100.000