

Sample Name : MOG(35-55)
 Cat No.: RP10245
 Lot No.: E7070419K
 Time Processed :20:38:24
 Month-Day-Year Processed :11/29/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.00	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment:ZJ19010324

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	11.197	2466	312	0.008
2	11.687	2934	417	0.010
3	11.854	7171	866	0.024
4	12.099	29024	3494	0.097
5	12.439	1290341	153338	4.300
6	12.790	28619152	3254229	95.370
7	14.087	5541	665	0.018
8	14.535	2150	216	0.007
9	14.992	33408	3306	0.111
10	16.417	1484	130	0.005
11	17.777	12521	962	0.042
12	18.822	1098	121	0.004
13	19.137	1190	154	0.004
Total		30008480	3418209	100.000