

Sample Name : OVA Peptide (257-264)
 Cat. No. : RP10611
 Lot. No. : PE4721612

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	15.526	16064	3116	0.309
2	16.407	3918	599	0.075
3	16.534	30923	5208	0.594
4	16.660	21766	3878	0.418
5	16.773	26078	4929	0.501
6	16.966	5021273	846524	96.483
7	17.200	3439	956	0.066
8	17.513	24667	4856	0.474
9	17.666	13995	2596	0.269
10	17.835	11878	1868	0.228
11	18.603	30330	3901	0.583
Total		5204331	878430	100.000