

Sample Name : CEF20  
 Cat No. : RP11781  
 Lot No. : E4031821F  
 Time Processed : 19:41:37  
 Month-Day-Year Processed : 06/18/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow: 1 ml/min  
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
10.00	Pumps	B.Conc	35
35.00	Pumps	B.Conc	95
37.00	Pumps	B.Conc	95
37.01	Pumps	B.Conc	5
45.00	Pumps	B.Conc	5
45.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	6.398	39002	6674	0.578
2	11.022	106762	6560	1.581
3	11.306	22024	4101	0.326
4	11.691	6415294	908553	95.021
5	11.938	16409	2694	0.243
6	12.291	74232	9905	1.099
7	20.071	77747	9999	1.152
Total		6751469	948487	100.000