

Sample Name : OVA Peptide (323-339)  
 Cat No. : RP10610  
 Lot No. : E2036620J  
 Time Processed : 8:04:16  
 Month-Day-Year Processed : 10/14/2020  
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow: 1 ml/min  
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment : ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	8.667	3519618	499333	98.783
2	9.077	7972	1620	0.224
3	11.968	21946	657	0.616
4	13.929	9421	1246	0.264
5	22.130	1055	91	0.030
6	24.705	2967	313	0.083
Total		3562979	503261	100.000