

Sample Name :Cys-TAT(47-57)
 Cat No. :RP20343
 Lot No. :E6787621E
 Time Processed :8:10:38
 Month-Day-Year Processed :05/25/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	0
5.00	Pumps	Pump A B.Conc	0
20.00	Pumps	Pump A B.Conc	30
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	0
35.00	Pumps	Pump A B.Conc	0
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	14.371	65720	18274	0.874
2	14.496	7318988	1222020	97.325
3	14.754	15882	10609	0.211
4	14.770	46587	11305	0.620
5	14.945	27864	3245	0.371
6	15.684	45119	5583	0.600
Total		7520161	1271036	100.000