

Sample Name: MOG(35-55)  
 Cat No.: RP10245  
 Lot No.:E5786119K  
 Time Processed :1:37:00 AM  
 Month-Day-Year Processed :11/16/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.00	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment:ZJ19010325

### <Chromatogram>

mV



### <Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	9.096	6857	276	0.122
2	10.078	2680	295	0.048
3	11.135	1194	113	0.021
4	11.424	3454	255	0.062
5	12.495	70100	1948	1.250
6	12.934	1852	469	0.033
7	14.085	1478	254	0.026
8	14.215	3556	488	0.063
9	14.308	1801	387	0.032
10	14.582	93457	12949	1.666
11	14.724	7374	2160	0.131
12	14.931	5386697	901992	96.022
13	15.565	1504	289	0.027
14	16.746	2273	242	0.041
15	18.144	2260	191	0.040

Peak#	Ret. Time	Area	Height	Area%
16	18.794	1348	143	0.024
17	19.336	1752	140	0.031
18	19.533	1078	125	0.019
19	20.258	1420	184	0.025
20	20.568	1863	204	0.033
21	21.650	2433	135	0.043
22	22.341	5351	517	0.095
23	23.664	8094	890	0.144
Total		5609875	924647	100.000