

Sample Name : MOG(35-55)
 Cat No. : RP10245
 Lot No. : E0655421A
 Time Processed :7:17:59 PM
 Month-Day-Year Processed :01/27/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)

Total Flow:1 ml/min

Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	5
35.00	Pumps	Pump A B.Conc	5
35.01	Controller	Stop	

<<Column Performance>>


<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm

Equipment: GK11010011

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	9.914	2272	223	0.021
2	13.607	10411	1717	0.098
3	14.854	4393	545	0.041
4	15.486	24559	2343	0.231
5	15.722	47248	5582	0.444
6	15.865	110491	10431	1.038
7	16.220	10284458	1325477	96.645
8	16.446	125229	11328	1.177
9	16.649	6397	987	0.060
10	17.293	1576	283	0.015
11	17.475	1263	195	0.012
12	20.516	4068	293	0.038
13	22.877	8544	527	0.080
14	23.076	9430	606	0.089
15	24.313	1188	171	0.011
Total		10641528	1360708	100.000