

Sample Name :OVA Peptide (323-339)

Cat. No. : RP10610

Lot. No. : E9540421B

Time Processed :2:47:21

Month-Day-Year Processed :03/03/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)

Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)

Total Flow:1 ml/min

Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	

<<Column Performance>>


<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm

Equipment: SS-CM-0309

### <Chromatogram>

mV



### <Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	8.184	6531673	872439	95.397
2	8.485	75602	5817	1.104
3	8.966	113866	11647	1.663
4	9.302	14665	1907	0.214
5	9.515	9562	1537	0.140
6	9.666	28261	2534	0.413
7	10.028	6622	957	0.097
8	10.383	9904	914	0.145
9	10.522	2767	453	0.040
10	10.737	7919	1052	0.116
11	10.899	22072	2616	0.322
12	11.530	11027	1213	0.161
13	11.890	1437	190	0.021
14	12.017	1670	277	0.024
15	12.144	4846	342	0.071
16	12.363	1243	258	0.018

Peak#	Ret. Time	Area	Height	Area%
17	12.480	1981	333	0.029
18	12.567	1687	344	0.025
Total		6846805	904831	100.000