

Sample Name : CEF20
 Cat. No. : RP11781
 Lot. No. : PE8041710

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	12.750	3323	643	0.203
2	12.973	2088	473	0.128
3	15.542	8921	1718	0.545
4	15.934	12360	2180	0.755
5	16.454	1601339	262369	97.831
6	16.718	3521	714	0.215
7	16.890	3327	682	0.203
8	17.153	1961	390	0.120
Total		1636841	269169	100.000