

Sample Name : CEF20  
 Cat No.: RP11781  
 Lot No.: E9701620F  
 Time Processed : 8:33:37 AM  
 Month-Day-Year Processed : 06/26/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow: 1 ml/min  
 Wavelength: 220 nm


<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.00	Controller	Stop	

<<Column Performance>>

<Detector A>  
 Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ19010325

<Chromatogram>



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	8.988	27937	4064	0.609
2	13.250	26208	3774	0.571
3	13.407	3934	650	0.086
4	13.695	35812	3783	0.780
5	14.002	29050	3127	0.633
6	14.291	21827	2558	0.475
7	14.496	4385777	569956	95.529
8	15.062	60497	5684	1.318
Total		4591043	593595	100.000