

Sample Information

Sample Name : RGD  
 Cat No : RP20297  
 Lot No : PE6571712

Pump A : 0.065% trifluoroacetic in 100% waters (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)


<<Pump>>  
 Flow : 1.0000 mL/min

<<Detector A>>  
 Wavelength Ch1 : 220 nm

<<LC Time Program>>

Time	Module	Command	Value	Comment
0.01	Pumps	Solvent B Conc.	5	
25.00	Pumps	Solvent B Conc.	65	
25.01	Pumps	Solvent B Conc.	95	
27.00	Pumps	Solvent B Conc.	95	
27.01	Pumps	Solvent B Conc.	5	
35.00	Pumps	Solvent B Conc.	5	
35.01	Controller	Stop		

<<Column Performance>>  
 <Detector A>  
 Column : Inertsil ODS-3 4.6 x 250 mm



Peak Table

Detector A Channel 1 220nm

Peak#	Ret. Time	Height	Area	Height%	Area%
1	7.385	49685	278466	1.652	1.113
2	7.895	30164	267984	1.003	1.072
3	8.301	2853578	23929939	94.880	95.685
4	8.579	21115	101183	0.702	0.405
5	8.813	53036	431614	1.763	1.726
Total		3007577	25009186	100.000	100.000