

Sample Name :Substance P  
 Cat. No. :RP10178  
 Lot. No. :E6036420K  
 Time Processed :7:37:55  
 Month-Day-Year Processed :11/13/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm  
 Equipment: SS-CM-0310

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	14.594	7760	894	0.117
2	14.938	3029	492	0.046
3	15.854	33263	1546	0.502
4	16.104	33013	5777	0.498
5	16.242	6445601	724514	97.323
6	16.504	55983	8535	0.845
7	16.874	44217	3852	0.668
Total		6622866	745609	100.000