

Sample Name :CEF20
 Cat No. :RP11781
 Lot No. :E8215221C
 Time Processed :6:46:28
 Month-Day-Year Processed :03/14/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>


Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	

<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010141

<Chromatogram>



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	10.340	1147	95	0.034
2	11.855	1900	179	0.057
3	12.159	1756	144	0.053
4	12.500	5955	486	0.179
5	12.624	5286	553	0.159
6	12.826	19144	2235	0.575
7	13.152	24660	2451	0.741
8	13.470	18443	1815	0.554
9	14.035	3193555	377896	95.942
10	14.689	34490	3376	1.036
11	15.200	3194	253	0.096
12	15.792	1142	133	0.034
13	16.175	1042	95	0.031
14	16.595	5020	447	0.151

Peak#	Ret. Time	Area	Height	Area%
15	16.925	1213	149	0.036
16	17.078	2162	213	0.065
17	17.608	1990	123	0.060
18	17.964	6529	368	0.196
Total		3328629	391011	100.000