

Sample Name :GRGDSPC
 Cat No. :RP20283
 Lot No. :PE7071905
 Time Processed :11:04:24
 Month-Day-Year Processed :05/12/19

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	5
31.00	Pumps	Pump A B.Conc	5
95.00	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ17010508

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	9.379	138853	8334	2.336
2	10.503	5754807	786422	96.834
3	10.808	29263	4073	0.492
4	13.216	20022	2735	0.337
Total		5942944	801565	100.000