

Sample Name:MOG(35-55)  
 Cat No. : RP10245  
 Lot No. : E6487221B  
 Time Processed :23:23:03  
 Month-Day-Year Processed :02/14/2021  
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ20010140

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	11.904	1691	394	0.057
2	12.046	24594	2646	0.830
3	12.197	9381	1620	0.317
4	12.408	102515	11422	3.459
5	12.703	2825805	367581	95.338
Total		2963986	383664	100.000