

Sample Name :Exendin-4  
 Cat No :RP10874  
 Lot No :PE0221903  
 Time Processed :9:44:31  
 Month-Day-Year  
 Processed :03/18/2019

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
33.00	Pumps	Solvent B Conc.	5
33.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment:ZJ17010508

### <Chromatogram>

mV



### <Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	20.025	35478	4240	0.461
2	20.879	7730	1368	0.100
3	21.488	238478	26913	3.099
4	21.743	7324835	551786	95.185
5	22.745	88840	10106	1.154
Total		7695360	594414	100.000