

Sample Information

Sample Name :RGD
 Cat. N.o :RP20297
 Lot. No. :PE7781804

Pump A : 0.065% trifluoroacetic in 100% waters (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)


<<Pump>>
 Flow : 1.0000 mL/min

<<Detector A>>
 Wavelength Ch1 : 220 nm

<<LC Time Program>>

Time	Module	Command	Value	Comment
0.01	Pumps	Solvent B Conc.	5	
25.00	Pumps	Solvent B Conc.	65	
25.01	Pumps	Solvent B Conc.	95	
27.00	Pumps	Solvent B Conc.	95	
27.01	Pumps	Solvent B Conc.	5	
35.00	Pumps	Solvent B Conc.	5	
35.01	Controller	Stop		

<<Column Performance>>
 <Detector A>
 Column : Inertsil ODS-3 4.6 x 250 mm



Peak Table

Peak#	Ret. Time	Height	Area	Height%	Area%
1	5.524	2082	12411	0.364	0.366
2	7.807	767	4492	0.134	0.133
3	8.352	1468	8512	0.256	0.251
4	8.621	1557	7623	0.272	0.225
5	8.688	3884	12904	0.678	0.381
6	8.827	559036	3322741	97.652	98.070
7	9.392	3682	19461	0.643	0.574
Total		572475	3388143	100.000	100.000