

Sample Name : OVA Peptide (257-264)
 Cat. No. : RP10611
 Lot. No. :PE3551611

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump B Conc.	5	
25.00	Pumps	Pump B Conc.	65	
25.01	Pumps	Pump B Conc.	95	
31.00	Pumps	Pump B Conc.	95	
31.01	Pumps	Pump B Conc.	5	
40.00	Pumps	Pump B Conc.	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	11.475	7082	995	0.201
2	11.893	1457	112	0.041
3	12.213	9205	1506	0.262
4	15.361	8609	1784	0.245
5	16.427	21648	3300	0.616
6	16.530	15152	2659	0.431
7	16.659	15919	2875	0.453
8	16.870	3382971	559129	96.221
9	17.458	20543	3090	0.584
10	17.559	9722	1948	0.277
11	17.785	7588	1171	0.216
12	18.068	15924	1912	0.453
Total		3515820	580481	100.000