

Sample Name :LCMV gp33-41
 Cat No: RP20257
 Lot No :E5082620J
 Time Processed :3:37:37 PM
 Month-Day-Year Processed :10/23/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

| Time | Module | Command | Value |
|-------|------------|---------|-------|
| 0.01 | Pumps | B.Conc | 5 |
| 25.00 | Pumps | B.Conc | 65 |
| 25.01 | Pumps | B.Conc | 95 |
| 31.00 | Pumps | B.Conc | 95 |
| 31.01 | Pumps | B.Conc | 5 |
| 40.00 | Pumps | B.Conc | 5 |
| 41.01 | Controller | Stop | |


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ19010140

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

| Peak# | Ret. Time | Area | Height | Area% |
|-------|-----------|----------|---------|---------|
| 1 | 10.045 | 56215 | 6635 | 0.434 |
| 2 | 10.379 | 189947 | 40124 | 1.468 |
| 3 | 10.515 | 12595491 | 1701837 | 97.343 |
| 4 | 10.792 | 97649 | 15274 | 0.755 |
| Total | | 12939302 | 1763870 | 100.000 |