

Sample Name: OVA Peptide (323-339)
 Cat No. : RP10610
 Lot No. : E3999920G
 Time Processed: 1:06:59 PM
 Month-Day-Year Processed :07/18/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

<<LC Time Program>>


Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
31.00	Pumps	Solvent B Conc.	95
31.01	Pumps	Solvent B Conc.	5
40.00	Pumps	Solvent B Conc.	5
40.01	Controller	Stop	

<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm
 Equipment: ZJ19010325

<Chromatogram>



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	13.667	77555	10834	0.397
2	14.023	125467	9307	0.643
3	14.469	199578	30494	1.023
4	14.672	18797097	1905160	96.336
5	14.900	122687	29151	0.629
6	15.017	139418	10612	0.715
7	15.408	50236	3015	0.257
Total		19512038	1998573	100.000