

Sample Name : MOG(35-55)  
 Cat.No : RP10245  
 Lot.No : E00586201  
 Time Processed: 21:22:53  
 Year-Month-Day: 2020/9/16

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow: 1 ml/min

<<Detector A>>  
 Wavelength Ch1 : 220 nm

<<LC Time Program>>


Time	Module	Command	Value	Comment
0.01	Pumps	Solvent B Conc.	5	
25.00	Pumps	Solvent B Conc.	65	
25.01	Pumps	Solvent B Conc.	95	
27.00	Pumps	Solvent B Conc.	95	
27.01	Pumps	Solvent B Conc.	5	
33.00	Pumps	Solvent B Conc.	5	
33.01	Controller	Stop		

<<Column Performance>>

<Detector A>  
 Column : Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ19010015

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	13.455	46731	4313	0.367
2	13.827	522214	64857	4.098
3	13.979	27927	7501	0.219
4	14.212	12134998	1676633	95.222
5	14.661	5200	973	0.041
6	14.806	6871	1150	0.054
Total		12743942	1755427	100.000