

Sample Name : Melittin,honey bee
 Cat No.: RP20415
 Lot No. : PE6631709
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm


Time	Unit	Command	Value	Comment
0.01	Pumps	Pump A B.Conc	5	
25.00	Pumps	Pump A B.Conc	65	
25.01	Pumps	Pump A B.Conc	95	
31.00	Pumps	Pump A B.Conc	95	
31.01	Pumps	Pump A B.Conc	5	
40.00	Pumps	Pump A B.Conc	5	
40.01	Controller	Stop		

<<Column Performance>>

<Detector A>

Column : AlltimaTM C18 4.6 x 250 mm

Chromatogram



1 Det.A Ch1 / 220nm

Peak Table

Detector A Ch1 220nm

Peak#	Ret. Time	Area	Height	Area %
1	13.492	28714	5308	0.703
2	15.941	16975	2437	0.416
3	16.704	9669	1417	0.237
4	17.479	10852	1647	0.266
5	19.883	26329	4923	0.645
6	19.936	61430	5526	1.504
7	20.319	49922	3674	1.222
8	20.760	3880750	312510	95.008
Total		4084641	337441	100.000