

Sample Name : LCMV gp33-41
 Cat. No. : RP20257
 Lot. No. : E1578321G
 Time Processed : 10:25:06 PM
 Month-Day-Year Processed : 07/17/2021
 Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	5
35.00	Pumps	Pump A B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment : GK11010011

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	13.164	142029	16444	1.465
2	13.854	9144	1562	0.094
3	14.130	2540	491	0.026
4	14.465	3675	407	0.038
5	14.941	9239366	1229563	95.322
6	15.476	296031	47600	3.054
Total		9692786	1296067	100.000