

Sample Name :Adrenomedullin (AM) (1-52), human
 Cat No :RP11288
 Lot No :E4732520K
 Time Processed :8:38:26 PM
 Month-Day-Year Processed :11/18/2020

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	5
35.00	Pumps	Pump A B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: SS-CM-0309

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	19.649	41023	2510	1.317
2	19.842	31064	2984	0.998
3	20.322	2974496	341944	95.521
4	20.923	55937	5010	1.796
5	21.349	11465	2054	0.368
Total		3113986	354503	100.000