

Sample Name :Exendin-4  
 Cat No. :RP10874  
 Lot No. :E7807221G  
 Time Processed :2:23:59 AM  
 Month-Day-Year Processed :07/30/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm  
 Equipment: ZJ19010141

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	8.102	1531	149	0.010
2	9.415	1789	125	0.012
3	10.805	1365	103	0.009
4	11.281	2329	160	0.015
5	12.054	33316	965	0.215
6	12.403	12552	1098	0.081
7	13.481	1365	206	0.009
8	14.210	1802	154	0.012
9	15.095	2289	230	0.015
10	15.643	4544	329	0.029
11	16.254	113832	6376	0.733
12	16.870	157019	9840	1.011
13	17.573	15068586	762868	97.044
14	18.746	73619	8729	0.474

Peak#	Ret. Time	Area	Height	Area%
15	18.892	42483	5609	0.274
16	22.274	2110	133	0.014
17	22.666	2937	151	0.019
18	23.235	1508	115	0.010
19	23.975	2535	169	0.016
Total		15527511	797509	100.000