

Sample Information

Sample Name :GRGDSPC  
 Cat. N.o :RP20283  
 Lot. No. :PE2261806

Pump A : 0.065% trifluoroacetic in 100% waters (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)


<<Pump>>  
 Flow : 1.0000 mL/min

<<Detector A>>  
 Wavelength Ch1 : 220 nm

<<LC Time Program>>

Time	Module	Command	Value	Comment
0.01	Pumps	Solvent B Conc.	5	
25.00	Pumps	Solvent B Conc.	65	
25.01	Pumps	Solvent B Conc.	95	
27.00	Pumps	Solvent B Conc.	95	
27.01	Pumps	Solvent B Conc.	5	
35.00	Pumps	Solvent B Conc.	5	
35.01	Controller	Stop		

<<Column Performance>>  
 <Detector A>  
 Column : Inertsil ODS-3 4.6 x 250 mm



Peak Table

Detector A Channel 1 220nm

Peak#	Ret. Time	Height	Area	Height%	Area%
1	5.405	8486	35574	0.910	0.574
2	5.477	8317	30798	0.892	0.497
3	6.094	11673	40760	1.252	0.658
4	6.552	7582	40169	0.813	0.648
5	6.690	860507	5899555	92.273	95.235
6	7.173	35999	147908	3.860	2.388
Total		932565	6194765	100.000	100.000