

Sample Name :Substance P
 Cat No :RP10178
 Lot No :E4386221E
 Time Processed :7:37:58 PM
 Month-Day-Year Processed :05/30/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	95
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm
 Equipment: GK11010011

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	8.874	1495	61	0.050
2	11.277	1431	125	0.048
3	11.925	66141	5521	2.213
4	12.465	15483	1029	0.518
5	13.331	2884355	439800	96.491
6	13.969	6195	1047	0.207
7	15.956	1208	158	0.040
8	17.954	1162	89	0.039
9	22.177	1916	175	0.064
10	22.531	2499	173	0.084
11	22.963	1468	69	0.049
12	23.568	1055	88	0.035
13	23.939	1259	122	0.042
14	24.310	1648	130	0.055
15	24.681	1945	189	0.065
Total		2989260	448777	100.000