

Sample Name:LCMV gp33-41  
 Cat No. : RP20257  
 Lot No. : E6927921G  
 Time Processed : 19:57:44  
 Month-Day-Year Processed : 07/09/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)  
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)  
 Total Flow:1 ml/min  
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump B Conc.	5
25.00	Pumps	Pump B Conc.	65
25.01	Pumps	Pump B Conc.	95
27.00	Pumps	Pump B Conc.	95
27.01	Pumps	Pump B Conc.	5
32.00	Pumps	Pump B Conc.	5
32.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column :Inertsil ODS-3 4.6 x 250 mm  
 Equipment:GK12010012

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	8.234	5011	766	0.063
2	9.132	2321	383	0.029
3	9.533	3284	424	0.041
4	9.863	58156	4617	0.734
5	10.100	195721	25503	2.469
6	10.999	40972	4497	0.517
7	11.217	2838	255	0.036
8	11.699	7533445	825507	95.044
9	12.317	1616	261	0.020
10	12.646	38870	4670	0.490
11	13.126	8553	763	0.108
12	13.452	35489	3457	0.448
Total		7926276	871102	100.000