

Sample Name :OVA Peptide (323-339)
 Cat No. :RP10610
 Lot No. :PE3711812
 Time Processed :20:27:47
 Month-Day-Year Processed :12/28/2018

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow:1 ml/min
 Wavelength:220 nm

<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Solvent B Conc.	5
25.00	Pumps	Solvent B Conc.	65
25.01	Pumps	Solvent B Conc.	95
27.00	Pumps	Solvent B Conc.	95
27.01	Pumps	Solvent B Conc.	5
35.00	Pumps	Solvent B Conc.	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Alltima™ C18 4.6 x 250 mm
 Equipment:ZJ17010508

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	9.778	6147	879	0.085
2	9.940	8002	989	0.111
3	10.374	7117251	1028681	98.696
4	10.620	60965	8045	0.845
5	10.860	18949	2364	0.263
Total		7211314	1040958	100.000