

Sample Name : OVA Peptide (323-339)
 Cat. No. : RP10610
 Lot. No. : E5384921E
 Time Processed : 8:31:09
 Month-Day-Year Processed : 05/17/2021

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
 Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
 Total Flow: 1 ml/min
 Wavelength: 220 nm

<<LC Time Program>>

| Time | Module | Command | Value |
|-------|------------|-----------------|-------|
| 0.01 | Pumps | Solvent B Conc. | 5 |
| 25.00 | Pumps | Solvent B Conc. | 65 |
| 25.01 | Pumps | Solvent B Conc. | 95 |
| 27.00 | Pumps | Solvent B Conc. | 95 |
| 27.01 | Pumps | Solvent B Conc. | 5 |
| 33.00 | Pumps | Solvent B Conc. | 5 |
| 33.01 | Controller | Stop | |


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-3 4.6 x 250 mm
 Equipment: ZJ17010507

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm

| Peak# | Ret. Time | Area | Height | Area% |
|-------|-----------|---------|--------|---------|
| 1 | 10.419 | 4850180 | 764697 | 97.395 |
| 2 | 10.588 | 59644 | 10679 | 1.198 |
| 3 | 10.696 | 27344 | 7042 | 0.549 |
| 4 | 10.788 | 32475 | 3286 | 0.652 |
| 5 | 11.211 | 1043 | 164 | 0.021 |
| 6 | 14.083 | 9230 | 878 | 0.185 |
| Total | | 4979916 | 786746 | 100.000 |